

भारतीय मानक

IS 3025 (Part 16) : 2023

Indian Standard

**जल और अपशिष्ट जल के नमूने लेने और
परीक्षण (भौतिक और रासायनिक) की
पद्धतियाँ**

**भाग 16 फिल्टर करने योग्य अवशेष (कुल घुले हुए
ठोस) 180 °C पर
(दूसरा पुनरीक्षण)**

**Methods of Sampling and Test
(Physical and Chemical) for Water
and Wastewater**

**Part 16 Filterable Residue (Total
Dissolved Solids) at 180 °C**

(Second Revision)

ICS 13.060.50

© BIS 2023



भारतीय मानक ब्यूरो
BUREAU OF INDIAN STANDARDS
मानक भवन, 9 बहादुर शाह ज़फर मार्ग, नई दिल्ली - 110002
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI - 110002
www.bis.gov.in www.standardsbis.in

January 2023

Price Group 4

Water Quality Sectional Committee, CHD 36

FOREWORD

‘This Indian Standard (Part 16) (Second Revision) was adopted by the Bureau of Indian Standards, after the draft finalized by the Water Quality Sectional Committee had been approved by the Chemical Division Council’.

Solid analyses are important for controlling water and wastewater treatment processes and assessing compliance with the regulatory requirements. Filterable residue is the term applied to the residue remaining in a weighed dish after the sample has been passed through a standard fiberglass filter and dried to constant mass at 103 °C – 105 °C or 179 °C – 181 °C.

The Technical Committee responsible for formulation of IS 3025 : 1964 ‘Methods of sampling and test (physical and chemical) for water used in industry’ decided to revise the standard and publish it in separate parts. This method superseded clause 12 of IS 3025 : 1964 ‘Methods of sampling and test (physical and chemical) for water used in industry’ and was one among the different parts published under IS 3025 series of standards. The first revision of this standard was published in 1984.

In this second revision the following modification have been incorporated:

- a) The gravimetric method has been updated; and
- b) The Amendment no.1 issued to the standard have been incorporated.

In the preparation of this standard, considerable assistance has been derived from the method no. 2540 C of — Standard Methods for the Examination of Water and Wastewater, published by the American Public Health Association, Washington, USA, 23rd Edition, 2017.

The composition of the committee responsible for the formulation of this standard is listed in Annex A.

In reporting the results of a test or analysis in accordance with this standard, if the final value observed or calculated, is to be rounded off, it shall be done in accordance with IS 2 : 2022 ‘Rules for rounding off numerical values (second revision)’.

Indian Standard

**METHODS OF SAMPLING AND TEST (PHYSICAL AND
CHEMICAL) FOR WATER AND WASTEWATER**

**PART 16 FILTERABLE RESIDUE (TOTAL DISSOLVED
SOLIDS) AT 180 °C**

(Second Revision)

1 SCOPE

1.1 This standard (Part 16) prescribes a gravimetric method for the determination of filterable residue in water and wastewater.

1.2 This method is applicable to all types of water and waste water.

2 REFERENCES

The following standards contain provisions which through reference in this text constitute provisions of this standard. At the time of publications, the editions indicated were valid. All standards are subject to revision and parties to agreements based on this standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below:

<i>IS No.</i>	<i>Title</i>
---------------	--------------

IS 1070	: Reagent grade water — Specification 1992
	(third revision)

3 SAMPLE HANDLING AND PRESERVATION

Preservation of the samples is not practical. Analysis should begin as soon as possible. Refrigeration or chilling to 4 °C to minimize microbiological decomposition of solids is recommended.

4 PRINCIPLE

A well-mixed sample is filtered through a standard glass-fiber filter and the filtrate is evaporated in a pre-weighed tared dish on steam-bath. The residue after evaporation is dried to constant mass in an oven at 180 °C ± 2 °C.

NOTE — The value of filterable residue obtained by drying at 180 °C conforms more closely to those obtained by summation of various constituents. (Since bicarbonates decompose to carbonates, only half of bicarbonate should be taken while summing up of the various constituents).

5 INTERFERENCES

The portion of total solids in a water sample that passes through a filter with a nominal pore size of 2.0 µm (or smaller) under specified conditions. Highly mineralized waters with a considerable calcium, chloride, magnesium, and/or sulphate content may be hygroscopic and require prolonged drying, proper desiccation and rapid weighing. Samples with high bicarbonate concentrations require careful, possibly prolonged drying at 180 °C ± 2 °C to ensure that bicarbonate is completely converted to carbonate.

6 APPARATUS

6.1 Dishes of 90 mm diameter with 100 ml capacity made of porcelain / platinum / high silica.

6.2 Wide-Bore Pippets — Class B glass, mechanical or electronic.

6.3 Graduated Cylinders — Class A.

6.4 Hot Plate/Block to Maintain <100 °C Temperature

6.5 Desiccator — Provided with a colour indicating desiccant.

6.6 Analytical Balance of 200g capacity and capable of weighing to 0.1 mg.

6.7 Magnetic Stirrer — With teflon coating stirring bars

6.8 Glass Fibre Filter Disc — (Whatman GF/C or equivalent) 22 mm to 125 mm dia, ≤ 2 µm nominal pore size without organic binder.

6.9 Oven — With thermostatic control for maintaining temperature up to 180 °C ± 2 °C.

6.10 Filtering Apparatus

Anyone of the following may be used:

6.10.1 Membrane Filter Funnel — Of various capacities to fit the selected filter.

IS 3025 (Part 16) : 2023

6.10.2 Gooch Crucible — 25 ml to 40 ml capacity, with Gooch-Crucible adapter.

6.11 Filtration apparatus with reservoir and coarse (40 µm - 60 µm) fritted disk with filter support

6.12 Suction Flask — Sufficient capacity for selected sample size.

7 PROCEDURE

7.1 Preparation of Glass Fibre Filter Disc

Place the glass fibre filter disk into the filtration apparatus with wrinkled surface up. While vacuum is applied, wash the dish with three successive 20 ml volumes of reagent grade water (*see* IS 1070). Remove all traces of water by continuing suction. 7.3 Preparation of Evaporating Dish

Heat the clean evaporating dish to 180 °C for 1 h. Cool in the desiccator, weigh and store in the desiccator until ready for use.

7.2 Filter a portion of the sample through any of the filters mentioned in **6.10**, Select volume of the sample which has residue between 25 mg and 250 mg or preferably between 100 mg to 200 mg. This volume may be estimated from values of specific conductance. To obtain a measurable residue. Successive aliquots of filtered sample may be added to the sample dish.

7.3 Stir volume of sample with a magnetic stirrer or shake it vigorously. Pipette this volume to a weighed evaporating dish placed on a steam-bath. Evaporation may also be performed in a drying oven. The temperature shall be lowered to approximately 98 °C to prevent boiling and splattering of the sample. After complete evaporation of water from the residue, transfer the

dish to an oven at 180 °C ± 2 °C and dry to constant mass, that is, till the difference in the successive weighing is less than 0.5 mg. Drying for a long duration (usually 1 h to 2 h) is done to eliminate necessity of checking for constant mass.

7.4 Using forceps, carefully remove filter from filtration apparatus and transfer to an inert weighing dish or pan as a support. If using a Gooch crucible, remove crucible and filter combination from the crucible adapter. Weigh the dish as soon as it has cooled avoiding residue to stay for long time as some residues are hygroscopic and may absorb water from desiccant that is not absolutely dry.

8 CALCULATION

Calculate the filterable residue from the following equation :

$$\text{Total Dissolved Solids, mg/l} = \frac{(A-B) \times 1000}{V}$$

where

A = final weight of dried residue + dish, in mg;

B = weight of dish, in mg; and

V = volume of sample, in ml.

9 REPORT

Report in whole numbers for less than 100 mg/l and to three significant figures for values above 100 mg/l. Report the temperature of determination.

10 PRECISION AND ACCURACY

The precision of the method is about 5 percent. Accuracy cannot be estimated because filterable residue as determined by this method is a quantity defined by the procedure followed.

ANNEX A
(Foreword)

COMMITTEE COMPOSITION

Water Quality Sectional Committee, CHD 36

<i>Organization</i>	<i>Representative(s)</i>
Chief Scientist, EPTRI, Hyderabad	SHRI N. RAVEENDHAR (Chairperson)
Andhra Pradesh Pollution Control Board, Vijaywada	SHRIMATI M. SREERANJANI SHRIMATI A. SRI SAMYUKTHA (<i>Alternate</i>)
Bhabha Atomic Research Centre, Mumbai	DR S. K. SAHU SHRI I. V. SARADHI (<i>Alternate</i>)
Central Institute of Mining and Fuel Research, Dhanbad	DR (MRS) BABLY PRASAD DR ABHAY KUMAR SINGH (<i>Alternate</i>)
Central Pollution Control Board, New Delhi	DR J. C. BABU
Confederation of Indian Industry, New Delhi	DR KAPIL K NARULA DR SIPIKA CHAUHAN (<i>Alternate</i>)
Delhi Jal Board, New Delhi	SHRI ASHUTOSH KAUSHIK SHRI SANJEEV KUMAR (YP) (<i>Alternate</i>)
Department of Civil Engineering, IIT Madras	DR LIGY PHILIP DR S. MATHAVA KUMAR (<i>Alternate</i>)
Envirocare Laboratories Pvt Ltd, Thane	DR PRITI AMRITKAR SHRI NILESH AMRITKAR (<i>Alternate</i>)
Drinking Water and Carbonated Beverages Sectional Committee, FAD 14, BIS	MEMBER SECRETARY
Gujarat Pollution Control Board, Gandhinagar	DR D. N. VANSADIA SHRI K. B. VAGHELA (<i>Alternate</i>)
Haryana State Pollution Control Board	SHRI JATINDER PAL SINGH
Himachal Pradesh State Pollution Control Board, Govt of Himachal Pradesh, Himachal Pradesh	DR T. B. SINGH ER PRAVEEN GUPTA
In personal capacity (1221, Mahatma Gandhi Road, P. O. — Haridevpur, Kolkata, 700082)	SHRI SANJIB KUMAR GOSWAMI
In Personal Capacity (S-168 A- Uppal Sothend, Sector 49, Sohna Road, Gurugram, 122018)	SHRI RAKESH MALHOTRA
In Personal Capacity, Hyderabad (H. No. 1-78/2/S/121/1, Sathi Reddy nagar colony, Boduppal, Hyderabad, Telangana, 500092)	SHRI N. MURALI MOHAN
Indian Agricultural Research Institute — Water Technology Centre, New Delhi	DR KHAJANCHI LAL DR RAVINDER KAUR (<i>Alternate</i>)
Indian Chemical Council, Mumbai	SHRI J. I. SEVAK DR MRITUNJAY CHAUBEY (<i>Alternate I</i>) DR N. D. GANGAL (<i>Alternate II</i>)

IS 3025 (Part 16) : 2023

<i>Organization</i>	<i>Representative(s)</i>
Indian Institute Of Chemical Technology, Hyderabad	DR SUNDERGOPAL SRIDHAR DR NIVEDITA SAHU (<i>Alternate</i>)
Indian Institute of Toxicology Research, Lucknow	DR S. C. BARMAN DR SATYAKAM PATNAIK (<i>Alternate</i>)
Indian Water Works Association	SHRI VIJAY CHARHATE
Karnataka State Pollution Control Board, Bengaluru	DR H. RUPADEVI DR GOURI GOLSANGI (<i>Alternate</i>)
Maharashtra State Pollution Control Board, Mumbai	DR V. R. THAKUR SHRI S. C. KOLLUR (<i>Alternate</i>)
National Environmental Engineering Research Institute, Nagpur	DR NOOR A. KHAN
National Institute of Oceanography, Vishakhapatnam	DR VVSS SARMA DR DURBAR RAY (<i>Alternate</i>)
NTPC Ltd, New Delhi	SHRI V. RAVI BABU SHRI SUDHIR DAHIYA (<i>Alternate</i>)
Shriram Institute for Industrial Research, New Delhi	DR VIVEK NARAYAN SINGH DR JAGDISH KUMAR (<i>Alternate</i>)
BIS Directorate General	SHRI AJAY KUMAR LAL, SCIENTIST 'E'/DIRECTOR AND HEAD (CHEMICAL DEPARTMENT) [REPRESENTING DIRECTOR GENERAL (<i>Ex-Officio</i>)]

Member Secretary
MS SHUBHANJALI UMRAO
SCIENTIST 'B'/ASSISTANT DIRECTOR
(CHEMICAL DEPARTMENT), BIS

This Page has been Intentionally left blank

Bureau of Indian Standards

BIS is a statutory institution established under the *Bureau of Indian Standards Act, 2016* to promote harmonious development of the activities of standardization, marking and quality certification of goods and attending to connected matters in the country.

Copyright

BIS has the copyright of all its publications. No part of these publications may be reproduced in any form without the prior permission in writing of BIS. This does not preclude the free use, in the course of implementing the standard, of necessary details, such as symbols and sizes, type or grade designations. Enquiries relating to copyright be addressed to the Head (Publication & Sales), BIS.

Review of Indian Standards

Amendments are issued to standards as the need arises on the basis of comments. Standards are also reviewed periodically; a standard along with amendments is reaffirmed when such review indicates that no changes are needed; if the review indicates that changes are needed, it is taken up for revision. Users of Indian Standards should ascertain that they are in possession of the latest amendments or edition by referring to the website- www.bis.gov.in or www.standardsbis.

This Indian Standard has been developed from Doc No.: CHD 36 (18540).

Amendments Issued Since Publication

Amend No.	Date of Issue	Text Affected

BUREAU OF INDIAN STANDARDS

Headquarters:

Manak Bhavan, 9 Bahadur Shah Zafar Marg, New Delhi 110002
Telephones: 2323 0131, 2323 3375, 2323 9402

Website: www.bis.gov.in

Regional Offices:

	Telephones
Central : 601/A, Konnectus Tower -1, 6 th Floor, DMRC Building, Bhavbhuti Marg, New Delhi 110002	{ 2323 7617
Eastern : 8 th Floor, Plot No 7/7 & 7/8, CP Block, Sector V, Salt Lake, Kolkata, West Bengal 700091	{ 2367 0012 2320 9474
Northern : Plot No. 4-A, Sector 27-B, Madhya Marg, Chandigarh 160019	{ 265 9930
Southern : C.I.T. Campus, IV Cross Road, Taramani, Chennai 600113	{ 2254 1442 2254 1216
Western : Plot No. E-9, Road No.-8, MIDC, Andheri (East), Mumbai 400093	{ 2821 8093

Branches : AHMEDABAD. BENGALURU. BHOPAL. BHUBANESHWAR. CHANDIGARH. CHENNAI.
COIMBATORE. DEHRADUN. DELHI. FARIDABAD. GHAZIABAD. GUWAHATI.
HIMACHAL PRADESH. HUBLI. HYDERABAD. JAIPUR. JAMMU & KASHMIR.
JAMSHEDPUR. KOCHI. KOLKATA. LUCKNOW. MADURAI. MUMBAI. NAGPUR.
NOIDA. PANIPAT. PATNA. PUNE. RAIPUR. RAJKOT. SURAT. VISAKHAPATNAM.